



Ref. Certif. No.

DE 3 - 40790

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**
CERTIFICAT D'ESSAI OC

Product

General Medical Devices
Vaporizer System

Name and address of the applicant

STORZ & BICKEL GMBH & CO. KG
In Grubenäcker 5-9
78532 Tuttlingen, GERMANY

Name and address of the manufacturer

STORZ & BICKEL GMBH & CO. KG
In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY

Name and address of the factory

STORZ & BICKEL GMBH & CO. KG
In Grubenäcker 5-9, 78532 Tuttlingen, GERMANY

Ratings and principal characteristics

Rated Input Voltage:	110- 120 V	220-240V
Rated Frequency:	50-60Hz	
Rated Input Power:	120VA	
Protection Class:	II	

Trade mark (if any)

vapormed

Model/type Ref.

Volcano Medic

Additional information (if necessary)

Certificate DE 3 - 40082 issued 2014-04-14 is replaced by this version due to technical changes.

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012
IEC 60601-1-6:2010
IEC 60601-1-6:2010/AMD1:2013
IEC 60601-1-11:2010as shown in the Test Report Ref. No.
which forms part of this certificate

713110770

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Date, 2017-07-19
CB 17 07 63179 061

Stefan Reis



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service